

IN THE SPECIFICATION

Please amend the paragraph beginning at page 10, line 6, as follows:

Although depending on the size of the treatment apparatus, weight of the tray and so forth, the dimensions of support beam 12 may be such that length L of the beam is 0.1 to 5 m, preferably 0.3 to 4 m, and more preferably 0.5 to 3 m. Width  $W_1$  of tray mounting portion 13 is 0.01 to 0.3 m, preferably 0.03 to 0.2 m, and more preferably 0.05 to 0.15 m. Furthermore, this does not apply to the case of the support beam and tray mounting portion being integrated into a single unit as described later. Width  $W_2$  of folded-back portion 14 may be 0.01 to 0.3 m, preferably 0.03 to 0.2 m, and more preferably 0.05 to 0.15 m. Although thickness can be selected arbitrarily, it is preferably 0.5 to 5 mm, and more preferably 1 to 3 mm. Folded-back portion 14 comprises inner surface 14' facing toward internal mounting portion 13 and outer surface 14'' facing away from internal mounting portion 13 and inner surface 14' is parallel to outer surface 14''. The length of the connecting portion may be 0.01 to 0.3 m, preferably 0.03 to 0.2 m, and more preferably 0.05 to 0.15 m.